




neodisher® PreStop



Corrosion inhibitor for the pre-treatment of surgical instruments

Spray foam

Fields of application:

- Corrosion inhibitor for spraying onto surgical instruments in operating theatres directly after use

Performance spectrum:

- Corrosion inhibitor: prevents pitting corrosion on stainless steel instruments which can be caused by adhesive residues of blood or physiological saline solution
- Self-acting pre-cleaning already in the OR
- Suitable for surgical instruments incl. MIS instruments
- Anodised aluminium must be tested for suitability before use

Special properties:

- Prevents drying of surgical residues by keeping the instruments wet
- Facilitates subsequent cleaning
- Inhibits the growth of micro-organisms on instruments during contact time (bacteriostatic)

Application and dosage:

neodisher PreStop is applied undiluted with the foam sprayer onto the surgical instruments within the disposal container directly after the operation; ensuring a complete wetting. The disposal container must then be kept closed. The instruments treated with neodisher PreStop can be stored for up to 72 hours. Then rinse carefully and reprocess as usual.

General instructions for use:

- For professional use only.
- For applying neodisher PreStop the foam sprayer which is within the scope of delivery is to be used.
- Do not mix with other products.

- neodisher PreStop must be rinsed off with water (preferably deionised).
- Reprocessing should comply with all ordinances pursuant to the regulations on medical devices and should be performed with appropriate validated processes.
- Please observe the reprocessing recommendations of the medical device manufacturers according to the requirements of the DIN EN ISO 17664.

Technical data:

pH-value	9.5 (undiluted, 20 °C)
Density	Approx.. 1.0 g/cm ³ (20 °C)

Ingredients:


< 5 % anionic and amphoteric surfactants also enzymes, preservatives (3-Iodo-2-propinyl butylcarbamate, 1,2-benzisothiazol-3(2H)-one)

CE-mark:

neodisher PreStop complies with European guidelines for medical devices.

If a serious incident occurs with the product, report it to the manufacturer and the relevant national authority.

Storage information:

Always store at a temperature between 0 °C and 25 °C. Keep container tightly closed. Keep away from sunlight. Usable for 2 years when stored as recommended. For expiry date refer to the stamp mark on the label behind the hourglass symbol .

Hazard and precautionary statements:

neodisher PreStop is not a hazardous product according to the CLP-Regulation (EC) No 1272/2008.

Dispose only when container is empty and closed. For disposal of product residues, refer to Safety Data Sheet.



neodisher[®] PreStop

For further safety information see safety data sheets. These are available e.g. at www.drweigert.com under the category "Service/Downloads".

MB 4086/2-4 Date of Issue: 05/2020



With the above information, to our current knowledge we describe our product regarding safety necessities, but we do not involve any quality description or promise certain properties.
